

# Casts of Sterilized/Disinfected Impressions, Long Term Evaluation

F. MARTINS<sup>1</sup>, J.A. REIS<sup>1</sup>, P. BRANCO<sup>1</sup>, A. FORJAZ<sup>1</sup>, J. PEREIRA<sup>1</sup>, P. MAURÍCIO<sup>1</sup>

<sup>1</sup> Oral Rehabilitation Clinic, Instituto Superior de Ciências da Saúde Egas Moniz, Monte da Caparica, Portugal

## INTRODUCTION

A successful rehabilitation depends on many factors such as dimensional stability, detail reproduction of impressions and models (Hamalian et al., 2011). Taking impressions is one of the crucial steps when it comes to oral rehabilitation. Impression quality determines the optimal adjustment of the restoration (Rupp et al., 2005; Balkenhol et al., 2010).

The aim of the impression material is to obtain a replica of the hard and soft tissues of the oral cavity in three dimensions and has to be dimensionally stable (Craig e Powers, 2002; Hamalian et al., 2011).

Nowdays, elastomers are considered the standard of care as the material for definitive impressions in Fixed Prosthodontics (Lee, 1999). Within the family of elastomers we can find polysulfides, condensation silicones, addition silicones and polyethers (Noort, 2007). The addition silicones and polyethers tend to be used most frequently due to it's physical and mechanical properties (Lee, 1999; Hamalian et al., 2011).

Disinfection procedures weren't used until the twentieth century. Impressions are contaminated by plaque, blood or saliva creating a vehicle of cross-infection for a variety of pathogens such as HIV, Hepatitis B, herpes, tuberculosis. Therefore it is necessary to control cross infection in clinical practice (Drennon e Johnson, 1990; Martin et al., 2007; Thomas et al., 2008; Rentzia et al., 2011).

The ADA Specification nº19 (1977) states that the maximum negative change in dimension 0,5% and the ISO 4823:2000 has a maximum of 1,5%.

## OBJECTIVES


Study the dimensional changes on gypsum casts poured with 2 addition silicones and a polyether, after the impressions were steam autoclave sterilized and stored.

## HYPOTHESIS

When subjected to autoclaving, the addition silicones and polyether suffer dimensional changes, resulting in casts with different dimensions of the matrix.

When subjected to autoclaving, the addition silicones and polyether don't suffer dimensional changes, resulting in casts with similar dimensions to the matrix.

## RESULTS AND DISCUSSION



The Impregum™ Penta™ Soft Polyether (3M ESPE™) casts had the most similar dimensions to the matrix with a maximum dimensional change of 0,15% after 2 weeks in B group and a maximum of 0,09% in A group after the same period. The Normosil Adición Putty Fast (Normon®) casts showed the most evident differences after 2 weeks with a maximum dimensional change of 0,29% and 0,34% in A and B group.

## CONCLUSIONS

The null hypothesis is rejected.

The gypsum casts shows dimensional changes of the impression materials after autoclaving.

The dimensional changes are below the maximum allowed by ADA Specification nº19 (1977) and ISO 4823:2000, therefore the steam autoclave sterilization should be considered a valuable disinfection procedure.

## MATERIAL AND METHODS



Impregum™ Penta™ Soft Polyether (3M ESPE™)



Aquasil Ultra Monophase (Dentsply®)




Normosil Adición Putty Fast (Normon®)



Pentamix 2 (3M ESPE™)



Metallic matrix side view



Metallic matrix top view




A group - 5 samples


B group - 5 samples

Chemical disinfection with sodium hypochlorite 5,25% - 10 min.

Chemical disinfection with sodium hypochlorite 5,25% - 10 min.  
+  
Steam autoclave sterilization at 134°C - 25 min.



Gypsum casts



a, b, c, ab and bc measured and averaged



Measurement example

## References

- ADA Specification nº19 (1977) for non-aqueous, elastomeric dental impression material.
- Balkenhol, M.; Haunschild, S.; Erbe, C. e Wöhrmann, B. (2010) "Influence of prolonged set time on permanent deformation of elastomeric impression materials." J Prosthet Dent., 103(5), pp. 288-294.
- Craig, R. G.; Powers, J. M. (2002) "Impression Materials - Chapter 12" in Restorative Dental Materials, 11a edição, pp. 329-390. Mosby, Texas.
- Drennon, G. e Johnson, G. H. (1990) "The effect of immersion disinfection of elastomeric impressions on the surface detail reproduction of improved gypsum casts." J Prosthet Dent., 63(2), pp. 233-241.
- Hamalian, T. A.; Nasr, E. e Chidiac, J. J. (2011) "Impression materials in fixed prosthodontics: influence of choice on clinical procedure." J Prosthodont., 20(2), pp. 153-160.
- Lee, E. A. (1999) "Predictable Elastomeric Impressions In Advanced Fixed Prosthodontics: A Comprehensive Review." Pract Periodontics Aesthet Dent., 11(4), pp. 497-504.
- Martin, N.; Martin, M. V. e Jedynakiewicz, N. M. (2007) "The dimensional stability of dental impression materials following immersion in disinfecting solutions." Dent Mater., 3, pp. 760-768.
- Noort, R. V. (2007) "Clinical Dental Materials" in Introduction to Dental Materials, 3a edição, Elsevier Limited, Reino Unido.
- Rentzia, A.; Coleman, D. C.; O'Donnell, M. J.; Dowling, A. H.; O'Sullivan, M. (2011) "Disinfection procedures: Their efficacy and effect on dimensional accuracy and surface quality of an irreversible hydrocolloid impression material." Journal of Dentistry, 38, pp. 133-140.
- Rupp, F.; Anmann, D.; Jacobi, A.; Groten, M.; Gels-Gestorfer, J. (2005) "Hydrophilicity of elastomeric non-aqueous impression materials during setting." Dental Materials, 21, pp. 94-102.
- Thomas, M.V.; Jarboe, G. e Frazer, R.G. (2008) "Infection control in the dental office." Dent Clin North Am, 52(3), pp. 609-28.